

CERTIFICATE OF CALIBRATION

Issued By Transmille Ltd.

Certificate Number EXAMPLE

Date of Issue 13 June 2007



Approved Signatory

Page 1 of 6 Pages



Transmille Ltd.
Unit 4, Select Business Centre
Lodge Road
Staplehurst, Kent. TN12 0QW.
TEL 01580 890700 FAX 01580 890711

Signatory 1 Signatory 2 Signatory 3

Customer : EXAMPLE

Date Received : 06 June 2007

Instrument :	System ID :	EXAMPLE	Job Number :	EXAMPLE
	Description :	Electrical Test Calibrator		
	Manufacturer :	Transmille		
	Model Number :	3200HGA		
	Serial Number :	EXAMPLE		
	Procedure Version :	3.10/N		

Environmental Conditions

Temperature :	20°C +/- 1°C	Mains Voltage :	240V +/- 12V
Relative Humidity :	50% +/- 20%	Mains Frequency :	50Hz +/- 1Hz

Comments

Instrument was allowed to stabilise for at least 12 hours before calibration.
Instrument calibrated with Bond lead supplied
4 Wire kelvin connections were used for ohms measurements below 10kOhms
Tests marked # are not UKAS accredited have been included for completeness

Calibration Information

The instrument was calibrated against laboratory standards whose values are traceable to recognised National Standards. The uncertainty limits quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Calibrated By : EXAMPLE

Date of Calibration : 13 June 2007

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to the units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0324
AFTER ADJUSTMENT RESULTS

Certificate Number
EXAMPLE

Page 2 of 6 Pages

Test Title	Applied Value	Reading	Uncertainties
Insulation Resistance			
10k Ω	10.000 0k Ω	10.000 0k Ω	430m Ω
20k Ω	20.000 0k Ω	20.001 2k Ω	610m Ω
30k Ω	30.000 0k Ω	29.995 0k Ω	730m Ω
40k Ω	40.000 0k Ω	40.002 2k Ω	860m Ω
60k Ω	60.000 0k Ω	59.981 9k Ω	1.1 Ω
100k Ω	100.000k Ω	99.965k Ω	2 Ω
200k Ω	200.000k Ω	199.945k Ω	4.1 Ω
400k Ω	400.000k Ω	399.830k Ω	7.6 Ω
600k Ω	600.000k Ω	599.808k Ω	11 Ω
1M Ω	1.000 00M Ω	1.000 06M Ω	22 Ω
2M Ω	2.000 00M Ω	1.999 77M Ω	170 Ω
3M Ω	3.000 00M Ω	2.999 77M Ω	250 Ω
4M Ω	4.000 00M Ω	4.001 51M Ω	330 Ω
5M Ω	5.000 00M Ω	5.001 75M Ω	410 Ω
6M Ω	6.000 00M Ω	5.986 08M Ω	500 Ω
7M Ω	7.000 00M Ω	6.985 86M Ω	580 Ω
8M Ω	8.000 00M Ω	7.985 75M Ω	660 Ω
9M Ω	9.000 00M Ω	8.987 56M Ω	740 Ω
10M Ω	10.000 0M Ω	9.972 5M Ω	830 Ω
20M Ω	20.000 0M Ω	19.948 4M Ω	13k Ω
30M Ω	30.000 0M Ω	29.923 9M Ω	19k Ω
40M Ω	40.000 0M Ω	39.887 4M Ω	25k Ω
50M Ω	50.000 0M Ω	50.019 2M Ω	32k Ω
60M Ω	60.000 0M Ω	59.969 0M Ω	38k Ω
70M Ω	70.000 0M Ω	69.967 3M Ω	44k Ω
80M Ω	80.000 0M Ω	79.926 1M Ω	51k Ω
90M Ω	90.000 0M Ω	89.874 8M Ω	57k Ω
100M Ω	100.000M Ω	99.786M Ω	64k Ω
200M Ω	200.000M Ω	199.252M Ω	1.6M Ω
400M Ω	400.000M Ω	398.933M Ω	3.2M Ω
600M Ω	600.00M Ω	603.39M Ω	4.9M Ω
800M Ω	800.00M Ω	803.81M Ω	6.5M Ω
1G Ω	1.000 0G Ω	0.994 1G Ω	8.1M Ω
2G Ω #	2.000G Ω	2.002G Ω	8.9M Ω
4G Ω #	4.00G Ω	4.01G Ω	18.1M Ω
6G Ω #	6.00G Ω	5.89G Ω	27M Ω
8G Ω #	8.00G Ω	7.90G Ω	36M Ω
10G Ω #	10.00G Ω	9.90G Ω	45M Ω

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0324
AFTER ADJUSTMENT RESULTS

Certificate Number
EXAMPLE

Page 3 of 6 Pages

Test Title	Applied Value	Reading	Uncertainties
------------	---------------	---------	---------------

Continuity Resistance

Connection to the 3200 insulation test terminals was made using 4 wire ohms with the system nulled when shorted at the terminals. The readings recorded are the resistance measured at the terminals and include any residual resistance of the 3200.

20m Ω	20.0m Ω	19.0m Ω	260 $\mu\Omega$
200m Ω	200.0m Ω	205.7m Ω	270 $\mu\Omega$
210m Ω	210.0m Ω	215.9m Ω	270 $\mu\Omega$
220m Ω	220.0m Ω	224.8m Ω	270 $\mu\Omega$
230m Ω	230.0m Ω	232.8m Ω	270 $\mu\Omega$
240m Ω	240.0m Ω	243.5m Ω	270 $\mu\Omega$
250m Ω	250.0m Ω	253.5m Ω	270 $\mu\Omega$
260m Ω	260.0m Ω	258.5m Ω	270 $\mu\Omega$
270m Ω	270.0m Ω	266.6m Ω	270 $\mu\Omega$
280m Ω	280.0m Ω	286.2m Ω	270 $\mu\Omega$
290m Ω	290.0m Ω	296.5m Ω	270 $\mu\Omega$
300m Ω	300.0m Ω	304.9m Ω	270 $\mu\Omega$
400m Ω	400.0m Ω	401.1m Ω	270 $\mu\Omega$
500m Ω	500.0m Ω	483.0m Ω	270 $\mu\Omega$
600m Ω	600.0m Ω	604.3m Ω	270 $\mu\Omega$
700m Ω	700.0m Ω	696.5m Ω	270 $\mu\Omega$
800m Ω	800.0m Ω	782.1m Ω	270 $\mu\Omega$
900m Ω	900.0m Ω	896.5m Ω	270 $\mu\Omega$
1 Ω	1.000 0 Ω	0.998 2 Ω	270 $\mu\Omega$
2 Ω	2.000 0 Ω	1.998 0 Ω	270 $\mu\Omega$
4 Ω	4.000 0 Ω	3.984 9 Ω	290 $\mu\Omega$
6 Ω	6.000 0 Ω	5.992 9 Ω	310 $\mu\Omega$
8 Ω	8.000 0 Ω	7.993 7 Ω	330 $\mu\Omega$
10 Ω	10.000 0 Ω	9.992 3 Ω	350 $\mu\Omega$
20 Ω	20.000 0 Ω	19.989 8 Ω	890 $\mu\Omega$
100 Ω	100.000 Ω	99.985 Ω	2.3m Ω
1k Ω	1.000 00k Ω	0.999 87k Ω	17m Ω

Continuity Current

100mA	100.0mA	100.0mA	120 μ A
200mA	200.0mA	200.0mA	120 μ A
300mA	300.0mA	300.4mA	240 μ A

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0324
AFTER ADJUSTMENT RESULTS

Certificate Number
EXAMPLE

Page 4 of 6 Pages

Test Title	Applied Value	Reading	Uncertainties
AC Voltage Output			
100 Volts Nom @ 50Hz	101.4V	101.4V	140mV
200 Volts Nom @ 50Hz	202.7V	202.6V	150mV
Line Volts @ 50Hz	240.1V	240.1V	160mV
300 Volts Nom @ 50Hz	304.1V	303.9V	190mV
400 Volts Nom @ 50Hz	405.1V	405.2V	230mV
Insulation Resistance Voltage Measurement			
50V	50.0V	50.0V	120mV
100V	100.0V	100.0V	120mV
250V	150.0V	149.9V	120mV
250V	200.0V	199.7V	120mV
250V	250.0V	250.0V	120mV
500V	500.0V	500.0V	120mV
1000V	1 000.0V	1 000.0V	120mV
Insulation Resistance Current Measurement			
0.500mA	0.500mA	0.499mA	4uA
1.000mA	1.000mA	1.000mA	4uA
Loop Resistance			
<i>Loop impedance was measured using 4 wire ohms connections between the earth pin of the 3200 loop test socket and the earth pin of the 3200 mains supply lead. The supply loop impedance was manually entered as zero and the measurement system was nulled. The recorded readings are the differences recorded from the zero value.</i>			
Loop Res.	0.055 7 Ω	0.055 7 Ω	1m Ω
Loop Res.	0.110 6 Ω	0.110 6 Ω	1.1m Ω
Loop Res.	0.226 8 Ω	0.226 5 Ω	1.1m Ω
Loop Res.	0.340 8 Ω	0.340 7 Ω	1.2m Ω
Loop Res.	0.504 7 Ω	0.504 7 Ω	1.3m Ω
Loop Res.	1.003 8 Ω	1.004 1 Ω	1.5m Ω
Loop Res.	5.085 1 Ω	5.084 8 Ω	3.5m Ω
Loop Res.	9.965 7 Ω	9.965 5 Ω	6m Ω
Loop Res.	100.355 Ω	100.355 Ω	18m Ω
Loop Res.	998.610 Ω	998.600 Ω	35m Ω
RCD Current			
10mA @ 50Hz	10.00mA	10.01mA	20uA
30mA @ 50Hz	30.00mA	30.00mA	80uA
100mA @ 50Hz	90.00mA	90.01mA	150uA
100mA @ 50Hz	100.00mA	100.06mA	160uA
100mA @ 50Hz	110.00mA	110.09mA	170uA
300mA @ 50Hz	300.0mA	299.9mA	840uA
1000mA @ 50Hz	1 000.0mA	1 001.2mA	1.7mA
2000mA @ 50Hz	2 000.0mA	2 000.3mA	3mA

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0324
AFTER ADJUSTMENT RESULTS

Certificate Number
EXAMPLE

Page 5 of 6 Pages

Test Title	Applied Value	Reading	Uncertainties
RCD Trip current			
150mA @ 300mA Range	150.0mA	149.8mA	240uA
RCD Trip time			
20ms	20.0ms	19.5ms	0.5ms
40ms	40.0ms	39.5ms	0.5ms
200ms	200.0ms	199.6ms	0.5ms
400ms	400.0ms	400.0ms	0.5ms
900ms	900.0ms	900.0ms	8.1ms
PAT : Insulation Resistance			
<i>The PAT Insulation Resistance is produced from the same decade resistance arm as used for the Insulation Resistance output. The following Tests are only to confirm the operation of the output switching. For the Full range of values, use the Insulation measurements on this certificate.</i>			
1M Ω	1.000 00M Ω	1.000 06M Ω	22 Ω
2M Ω	2.000 00M Ω	1.999 63M Ω	170 Ω
4M Ω	4.000 00M Ω	4.001 39M Ω	330 Ω
6M Ω	6.000 00M Ω	5.985 73M Ω	500 Ω
8M Ω	8.000 00M Ω	7.985 24M Ω	660 Ω
10M Ω	10.000 0M Ω	9.971 6M Ω	830 Ω
PAT : Earth Bond Resistance			
<i>The resistances recorded include the resistance of the PAT test mains lead (approx 25milliohms)</i>			
PAT Lead No	1 000a	1 732a	
PAT Lead Resistance	25.0m Ω	23.8m Ω	260u Ω
Bond Res.	0.044 9 Ω	0.045 0 Ω	1m Ω
Bond Res.	0.103 6 Ω	0.103 6 Ω	1m Ω
Bond Res.	0.157 8 Ω	0.157 9 Ω	1.1m Ω
Bond Res.	0.274 4 Ω	0.274 4 Ω	1.1m Ω
Bond Res.	0.388 7 Ω	0.388 6 Ω	1.2m Ω
Bond Res.	0.552 6 Ω	0.552 6 Ω	1.3m Ω
Bond Res.	1.052 0 Ω	1.052 0 Ω	1.5m Ω
Bond Res.	5.132 0 Ω	5.133 0 Ω	3.5m Ω
Bond Res.	10.015 6 Ω	10.015 7 Ω	6m Ω
Bond Res.	100.425 Ω	100.426 Ω	18m Ω
Bond Res.	1 002.130 Ω	1 002.118 Ω	35m Ω
PAT : Earth Bond Current			
100mA @ 50Hz	0.100A	0.100A	1.2mA
10A @ 50Hz	8.00A	8.00A	15mA
10A @ 50Hz	10.00A	10.00A	16mA
20A @ 50Hz	20.00A	20.01A	22mA
PAT: LOAD TESTS			
S/C TEST	0.00 Ω	0.39 Ω	10m Ω
O/C TEST	---	Pass	
0.13KVA TEST	440.0 Ω	439.4 Ω	0.1 Ω
PAT: LEAKAGE CURRENT TEST			
Leakage @ 240V	2.000mA	2.000mA	9.3uA
Leakage @ 240V	4.700mA	4.710mA	9.3uA

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0324
AFTER ADJUSTMENT RESULTS

Certificate Number
EXAMPLE

Page 6 of 6 Pages

Test Title	Applied Value	Reading	Uncertainties
Leakage @ 240V	7.700mA	7.720mA	9.3uA

End of results